

CS Disposable Lateral Head Support (Adult)



It is recommended that the patient's head is turned every 2 to 4 hours per ICS guidelines

No need to disconnect the tubing Single use to aid infection control





Manufacturer

UK: Care Surgical Ltd Unit 6 Ringtail Road Burscough, L40 8JY Tel: 01704 336671

USA: Care Surgical LLC 43 Broad Street Suite B106, Hudson, MA 01749, USA Tel: 866-243-4107





Assessed

Medical Device UK Conformity Do not reuse CE Mark











LOT:



www.care-surgical.com




Product Code: CSM-2410 Rev, 6 01/24



Head-to-Toe Check list

This check should be completed once the patient is in the prone position and before surgical drapes are applied. At all stages, check that no tubing is trapped between the patient and the supports, and that all bony protrusions are padded.




+ Chin and throat not touching the chest support

+ Natural, neutral neckline

Chest, abdomen and pelvis

- + ECG electrodes positioned on the patient's back
- + Breast positioned medically with minium pressure
- + Adequate space to avoid compression of the abdomen
- + Male genitalia free of the pelvic support



The arms

- + Arms forward and not abducted beyond 90° or arms secure by the patients side
- + No pressure on the axillary nerve
- + No body weight supported on the arms

Legs and feet

- + No presssure on the thighs
- + Ankles raised

To see our full range of Care Surgical Medical Products, visit www.care-surgical.com



Manufacturer

UK: Care Surgical Ltd USA: Care Surgical LLC Unit 6 Ringtail Road Burscough, L40 8JY Tel: 01704 336671

43 Broad Street Suite B106, Hudson, MA 01749, USA Tel: 866-243-4107



Medical Device UK Conformity Do not reuse

Assessed





CE Mark











www.care-surgical.com



Care Surgical Supplies Limited, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland

Product Code: CSM-2410 Rev.6 01/24