
CARE SURGICAL CS Hip Supports (pair)

Patient Supports for use on the CS Spine Frame or spinal operating tables



Refer to the label on the support to ensure it is placed in the correct position on the table.



Manufacturer

UK: Care Surgical Ltd Unit 6 Ringtail Road Burscough, L40 8JY Tel: 01704 336671

USA: Care Surgical LLC 43 Broad Street Suite B106, Hudson, MA 01749, USA Tel: 866-243-4107

www.care-surgical.com



EP Care Surgical Supplies Limited, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland





Product Code: CSM-2717 Rev.7 10/23




Latex Free Non Sterile UK Conformity


NON STERIL

> Conformity Cl Assessed


CE Mark Maximum Weight



Place the supports on the frame so that the brackets are over the bars



Snap the locking clip shut to 2 secure the support in position



To adjust the position of the supports, release the 3 clip(s), raise the patient's chest or hips, slide the supports to the correct position, lock the clip(s)

Ensure the iliac crest is in the centre of the hip pads.

These products will support a patient weight of up to 250kg. Do not exceed the weight capacity of the operating table.

Cleaning instructions:



Clean using standard hospital wipes

Do not immerse in water

Do not autoclave

Manufacturer

UK: Care Surgical Ltd Unit 6 Ringtail Road Burscough, L40 8JY Tel: 01704 336671

USA: Care Surgical LLC 43 Broad Street Suite B106, Hudson, MA 01749, USA Tel: 866-243-4107

www.care-surgical.com

Care Surgical Supplies Limited, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland EC REP



Product Code: CSM-2717 Rev.7 10/23

Medical Device Reuse Latex Free Non Sterile UK Conformity

STERIL

Assessed

CE Mark Weight

Maximum