



Technical Data Sheet

CSM-2750 CS Standard Replacement Stirrup Pads (Pair)



Description

- Reusable - clean using non-alcohol hospital wipes
- Secures and cushions patient's feet and legs in lithotomy stirrups
- Compatible with the Allen® Medical OEM Stirrups
- For use in a variety of surgical procedures
- Reduces the incidence of pressure ulcers and marking
- Made from supportive memory foam, fully encapsulating the patient's foot
- Maximum patient weight: 350lbs (159kg)

Application Area

- OR
- Anaesthesia

Measurements

- Length 570mm/22.4"
- Width 560mm/22"
- Height 100mm/3.9"

Material Composition

- Polyurethane foam
- Latex & phthalate free
- Transfer coated PU with anti-bacterial treatment on knitted polyester

Cleaning

- Clean using non-alcohol hospital wipes

Order Information

Product Code	Product Name	Qty
CSM-2750	CS Standard Replacement Stirrup Pads (Pair)	1 Pair

Packaging Information

- Each pair is packaged in a recyclable polyethylene bag
- 1 pair is packaged into a cardboard box for sale

Storage Instructions

- First in first out
- Store in a cool dry place
- Avoid direct sunlight
- Do not store anything heavy on top
- Avoid storing where sharp objects could pierce the skin of the product

Contraindications

Incorrect positioning of the patient on this product can cause:

- Pressure marking or ulcers
- Neurological damage, which can cause paraplegia
- Venous pressure, which can lead to ischaemia

Examine the pad for damage before and after each use. If the skin is pierced or torn, it cannot be cleaned effectively and should be replaced.

- Do not autoclave
- Only to be used by trained medical personnel

Further Information

- Disposable/Reusable: Reusable
- Sterility: N
- CE/FDA/TGA/UKCA:Y
- MD Class: 1
- Basic UDI/GTIN No: 5060624123689

Labelling



Manufacturer

UK: Care Surgical Ltd
Unit 6 Ringtail Road
Burscough, L40 8JY
United Kingdom
Tel: +44 1704 336671

US: Care Surgical LLC
Marlborough
MA 01752
United States
Tel: +1-866-243-4107

www.care-surgical.com

EC REP

Care Surgical Supplies Ltd, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland



Medical Device



Reuse



Latex Free



Non Sterile



UK Conformity Assessed



CE Mark

Rev.01 04/03/2024