

#### Description

- Adaptor for use on standard dual-rail tables, with EU table rail. It is used to provide a secure method of mounting (holding) other medical devices, such armboards
- Reusable clean using alcohol-based hospital wipes
- For use with dual rail tables, compatible with CS Carbon Spine Extension Frame

## **Application Area**

- OR
- Anaesthesia

#### Measurements

- Length 155mm/6.10"
- Width 85mm/3.35"
- Height 70mm/2.76"

#### **Material Composition**

- Aluminium,
- Stainless steel
- Plastic
- Latex & phthalate free

### Cleaning

· Clean using alcohol-based hospital wipes

#### **Order Information**

Product Code	Product Name	Qty
CSM-2739-1	CS Table Rail Adaptor EU	1 Unit

## **Packaging Information**

- Each unit is packaged in a recyclable polyethene bag
- 1 unit is packaged into a cardboard box for sale

# **Storage Instructions**

- First in first out
- · Store in a cool dry place

#### Contraindications

- Incorrect attachement to dual rail tables or the CS Carbon Spine Extension Frame could result in the attachments failing
- Do not use if the adaptor appears to be defective or damaged
- Only to be be used by trained medical personnel
- Do not exceed the device safe working load
- Do not immerse in water
- Do not autoclave
- After each use, clean the device with alcohol-based wipes
- Make sure that the device is dry before you store it or use it again

#### **Further Information**

- Disposable/Reusable: Reusable
- · Sterility: N
- CE/FDA/TGA/UKCA:Y
- MD Class: 1
- Basic UDI/GTIN No: 5060624123573

# Labelling



UK: Care Surgical Ltd Unit 6 Ringtail Road Burscough, L40 8JY United Kingdom Tel: +44 1704 336671

US: Care Surgical LLC Marlborough MA 01752 United States Tel: +1-866-243-4107

Manufacturer
















EC REP

Care Surgical Supplies Ltd, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland

Australian Sponsor, Summit Surgical PO Box 60, Landsborough QLD 4550 Australia