



Technical Data Sheet

CSM-2733-2 CS Armboard - 3"



Description

- 3"/80mm Armboard Pad
- For use on standard compatible armboards to support patients' arms
- Reusable
- Suitable to support patients' arms in a wide range of positions
- Number of uses: Unrestricted, subject to damage (see damage indicators in the Contraindications column)

Application Area

- OR
- Anaesthesia

Measurements

- Length 670mm/26.38"
- Width 140mm/5.51"
- Height 80mm/3.15"

Material Composition

- Polyurethane foam
- Latex & phthalate free
- Transfer coated PU with anti-bacterial treatment on knitted polyester

Cleaning

- Clean using non-alcohol hospital wipes

Order Information

Product Code	Product Name	Qty
CSM-2733-2	CS Armboard - 3"	1 Unit

Packaging Information

- Each unit is packaged in a recyclable polyethylene bag
- 1 unit is packaged into a cardboard box for sale

Storage Instructions

- First in first out
- Store in a cool dry place
- Avoid direct sunlight
- Do not store anything heavy on top
- Avoid storing where sharp objects could pierce the skin of the product

Contraindications

Incorrect positioning of the patient on this product can cause:

- Pressure marking or ulcers
- Neurological damage, which can cause paraplegia
- Venous pressure, which can lead to ischaemia

Examine the pad for damage before and after each use. If the skin is pierced or torn, it cannot be cleaned effectively and should be replaced.

- Do not immerse in water
- Do not autoclave
- Only to be used by trained medical personnel

Further Information

- Disposable/Reusable: Reusable
- Sterility: N
- CE/FDA/TGA/UKCA: Y
- MD Class: 1
- Basic UDI/GTIN No: 5060624121982

Labelling



Manufacturer

UK: Care Surgical Ltd
Unit 6 Ringtail Road
Burscough, L40 8JY
United Kingdom
Tel: +44 1704 336671

US: Care Surgical LLC
Marlborough
MA 01752
United States
Tel: +1-866-243-4107

www.care-surgical.com



Care Surgical Supplies Ltd, Office 2, 12a Lower Main
Street, Lucan Co. Dublin, K78 X5P8 Ireland



Medical Device



Reuse



Latex Free



Non Sterile



UK Conformity
Assessed



CE Mark

Rev.02 04/03/2024