Technical Data Sheet

CSM2762 CS Blade Clamp Plus - US

Description

- Table rail attachment for US, to be used with Blade type accessories
- Reusable - clean using alcohol-based hospital wipes
- For use on compatible surgical tables in surgical environments and ICUs
- Intended to fix or mount Kyra style stirrups or other blade-style instrumentation to operating table rails to the device

Application Area

- OR

Measurements

- Length 75mm/3"
- Width 50mm/2"
- Height 75mm/3"

Material Composition

- Latex & phthalate free
- Main Body: - Aluminium 6061
- Screw - Stainless Steel 316
- Washer - Nylon
- Handle - Stainless steel 304 (Insert) and Glass-Fibre Reinforce Polyamide (PA)

Order Information

Product Code	Product Name	Qty
CSM2762	CS Blade Clamp Plus - US	1 Unit

Packaging Information

- Each unit is packaged in a recyclable polyethene bag
- 1 unit is packaged into a cardboard box for sale

Storage Instructions

- First in first out
- Store in a cool dry place
- Avoid direct sunlight
- Transport and store in dry conditions

Contraindications

- Incorrect attachment to the surgical tables could result in the attachments failing
- Do not use if the blade clamp appears to be defective or damaged
- Only to be be used by trained medical personnel
- Do not exceed the device safe working load
- Do not immerse in water
- Do not autoclave
- After each use, clean the device with alcohol-based wipes
- Make sure that the device is dry before you store it or use it again

Further Information

- Disposable/Reusable: Reusable
- Sterility: N
- CE/FDA/TGA/UKCA:Y
- MD Class: 1
- Basic UDI/GTIN No: 5060624124013

Labelling

Manufacturer

UK: Care Surgical Ltd
Unit 6 Ringtail Road
Burscough, L40 8JY
United Kingdom
Tel: +44 1704 336671

US: Care Surgical LLC
43 Broad Street
Suite B106, Hudson
MA 01749, USA
Tel: +1-866-243-4107

www.care-surgical.com

EU REP

Care Surgical Supplies Ltd, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland

Medical Device · Reuse · Latex Free · Non Sterile · UK Conformity Assessed · CE Mark

Rev.02 01/04/2025