



Technical Data Sheet

CSM3048 CS Mayfield / Doro Adaptors (pair)



Description

- Accessory for CS Universal Spine Extension Frame (CSM2957)
- Bolt to CSM2957 to allow a cranial stabilisation device to be attached
- For use in surgical environments and ICUs
- Reusable - clean using alcohol-based hospital wipes

Application Area

- OR
- Anaesthesia

Measurements

- Length: 90mm/3.54"
- Width: 41mm/1.61"
- Height: 65mm/2.56"

Material Composition

- Technopolymer
- Zinc-plated steel threaded stud
- Aluminium
- Stainless Steel

Order Information

Product Code	Product Name	Qty
CSM3048	CS Mayfield / Doro Adaptors (pair)	2 Units

Packaging Information

- Each unit is packaged in a recyclable polyethene bag
- 2 units are packaged into a cardboard box for sale

Storage Instructions

- First in first out
- Store in a cool dry place
- Avoid direct sunlight
- Transport and store in dry conditions

Contraindications

- Incorrect attachment to the CSM2957 could result in the attachments failing
- Do not use if the adaptors appear to be defective or damaged
- Only to be used by trained medical personnel
- Do not exceed the device safe working load
- Do not immerse in water
- Do not autoclave
- After each use, clean the device with alcohol-based wipes
- Make sure that the device is dry before you store it or use it again

Further Information

- Disposable/Reusable: Reusable
- Sterility: N
- CE/FDA/UKCA: Y
- MD Class: 1
- GTIN No: 5060624126987/5060624127007

Labelling



Manufacturer

UK: Care Surgical Ltd
Unit 6 Ringtail Road
Burscough, L40 8JY
United Kingdom
Tel: +44 1704 336671

US: Care Surgical LLC
43 Broad Street
Suite B106, Hudson
MA 01749, USA
Tel: +1-866-243-4107

www.care-surgical.com



Care Surgical Supplies Ltd, Office 2, 12a Lower Main Street, Lucan Co. Dublin, K78 X5P8 Ireland



Medical Device



Reuse



Latex Free



Non Sterile



UK Conformity Assessed



CE Mark



Max Patient Weight